
49026 Milmont Drive, Fremont, CA 94538 Tel: +1 (800) 421-4149, Fax: +1 (510) 824-1490, support@biogenex.com



Catalog No.	Description
AN747-5M	6 ml of Ready-to-Use Antibody for Use with BioGenex Super
	Sensitive Detection Systems or Other Equivalent Detection
	Systems
AN747-10M	10 ml of Ready-to-Use Antibody for Use with BioGenex
	Super Sensitive Detection Systems and BioGenex Automated
	Staining Systems
NU747-UC	1 ml of Concentrated Antibody for Use with BioGenex Super
	Sensitive Detection Systems or Other Equivalent Detection
	Systems
NU747-5UC	0.5 ml of Concentrated Antibody for Use with BioGenex
	Super Sensitive Detection Systems or Other Equivalent
	Detection Systems
AY747-YCD	Ready-to-Use Antibody in RFID tagged vial for use on the
	Xmatrx [®] Elite/Ultra Staining System, 200 tests
AY747-50D	Ready-to-Use Antibody in RFID tagged vial for use on the
	Xmatrx [®] Elite/Ultra Staining System, 50 tests

Immunogen	Clone	Species	Ig Class
Recombinant full length mouse calretinin	SP13	Rabbit	IgG
protein			

Intended Use

This antibody is currently available for in vitro diagnostic use. This antibody is designed for the specific localization of Calretinin in formalinfixed, paraffin-embedded tissue sections.

Summary and Explanation

This antibody recognizes a protein of 31.5kDa, identified as Calretinin. Calretinin is an intracellular calcium-binding protein belonging to the troponin C superfamily characterized by a structural motif described as the EF-hand domain. It is abundantly expressed in central and peripheral neural tissues, particularly in the retina and in the neurons of the sensory pathways, and calretinin may play an important role in the survival of nerve cells during disturbances in calcium homeostasis. Calretinin is also expressed by both normal and neoplastic mesothelial cells, and it has been suggested as a useful marker for the identification of malignant mesotheliomas of the epithelial type and for the differentiation of these malignancies of lung adenocarcinoma.

Principles of the Procedure

The demonstration of antigens by immunohistochemistry (IHC) is a twostep process involving first, the binding of a primary antibody to the antigen of interest, and second, the detection of bound antibody by a chromogen. The primary antibody may be used in IHC using manual techniques or using BioGenex Automated Staining System. BioGenex offers a variety of Super Sensitive detection systems including Link-Label and Polymer-based technologies to detect the chromogenic signal from the stained tissues and cells.

Positive and negative controls should be run simultaneously with all patient specimens. If unexpected staining is observed which cannot be explained by variations in laboratory procedures and a problem with the antibody is suspected, contact BioGenex Technical Support at 1.800.421.4149 or your local distributor.

Reagents Provided

Rabbit Monoclonal Antibody to Calretininis affinity purified and diluted in PBS, pH 7.2, containing 1% BSA and 0.09% sodium azide.



Dilution of Primary Antibody

BioGenex Ready-to-Use antibodies have been optimized for use with the recommended BioGenex Detection System and should not require further dilution. Further dilution may result in loss of sensitivity and must be validated by the user.

BioGenex Concentrated antibodies must be diluted in accordance with the staining procedure when used with the recommended BioGenex Detection System (see staining procedure section below). Use of any detection methods other than the recommended systems and protocols require validation by the user.

Materials Required But Not Provided

All the reagents and materials required for IHC are not provided. Pretreatment reagents, Super Sensitive detection systems, control slides, control reagents and other ancillary reagents are available from BioGenex. Please refer to the product insert(s) of the BioGenex Super Sensitive IHC detection systems for detailed protocols and instructions.

The IHC procedure may need other lab equipment that is not provided including oven or incubator (capable of maintaining 56-60°C), BioGenex Automated Staining System, Humidity Chamber, Microwave oven, Staining Jars or baths, Timer (capable of 3-20 minute intervals), Wash Bottles, Absorbent Wipes, Microscope slides (pre-treated with poly-L-Lysine), Coverslips, Lens paper and Light microscope with magnification of 200X.

Storage and Handling

Antibodies should be stored at 2-8°C without further dilution. Fresh dilutions, if required, should be made prior to use and are stable for up to one day at room temperature (20-26°C). Unused portions of antibody preparations should be discarded after one day to minimize microbial contamination and increase in nonspecific staining. Diluted antibody preparations can be refrigerated or frozen for extended shelf life.

This antibody is suitable for use until expiry date when stored at 2-8°C. Do not use product after the expiration date printed on vial. If reagents are stored under a condition other than those specified in the package insert, they must be verified by the user (U.S. Congress, 1992).

The presence of precipitate or an unusual odor indicates that the antibody is deteriorating and should not be used.

Specimen Collection and Preparation

Tissues fixed in 10% (v/v) formalin, prior to paraffin embedding, are suitable for use. For further details on specimen preparation please refer to -Kiernan, 1981: Sheehan & Hrapchak, 1980.

The user is advised to validate the use of the products with their tissue specimens prepared and handled in accordance with their laboratory practices.

Treatment of Tissues Prior to Staining

Pretreatment of tissues, if any, should be done as suggested in the staining procedure section below.

Precautions

This antibody contains no hazardous material at a reportable concentration according to U.S. 29 CFR 1910.1200, OSHA Hazard Communication Standard and EC Directive 91/155/EC. However, this product contains sodium azide, at concentrations of less than 0.1%. Sodium azide is not classified as a hazardous chemical at the product concentrations. However, toxicity information regarding sodium azide at product concentrations has

Doc. No. 932-747N-EN Rev. D Release Date: 18-Jul-2016

49026 Milmont Drive, Fremont, CA 94538 Tel: +1 (800) 421-4149, Fax: +1 (510) 824-1490,

support@biogenex.com

not been thoroughly investigated. Sodium azide may react with lead or copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing (Center for Disease Control, 1976, National Institute for Occupational Safety and Health, 1976). For more information, a Safety Data Sheet (SDS) for sodium azide in pure form is available upon request. Dispose unused reagents according to local, State and Federal regulations. Wear suitable Personal Protective Equipment, do not pipette reagents by mouth, and avoid contact of reagents and specimens with skin and mucous membranes. If reagents or specimens come in contact with sensitive area, wash with copious amounts of water.

Refer to appropriate product inserts for instructions of use and safety information on detection reagents and other materials, which may be used with the antibody.

Recommended Protocol

Refer to the following table for conditions specifically recommended for this antibody. Refer to the detection system package insert for guidance on specific staining protocols or other requirements.

Parameter	BioGenex Recommendations	
Control Tissue	MESOTHELIOMA as available with	
	Biogenex FB-747N* & FG-747N*	
Recommended Dilution for	100 in HK941-YAK	
Concentrated Ab		
Recommended Pretreatment	EZ-AR2 Elegance (HK547-XAK)	
(manual/i6000)**		
Recommended Pretreatment(Xmatrx)	EZ-AR2 Elegance (HX032-YCD)	
Antibody incubation(manual/i6000)	30 mins at RT	
Antibody incubation (Xmatrx)	45 mins at 25°C	
Detection System for manual, Xmatrx	Use two-Step Super Sensitive™	
& i6000 -Open systems***	Polymer-HRP IHC Detection	
	System/DAB available from BioGenex	
	(QD400 for Manual and QD410 for	
	Automation).	
	or	
	For One-Step use Super Sensitive TM	
	Polymer-HRP Detection Kit/DAB,	
	available from BioGenex (QD620 or	
	QD630)	

^{*}FB: positive control barrier slides, FG: positive control non barrier slides. Xmatrx requires barrier slides.

Quality Control

The recommended positive control tissue for this antibody is MESOTHELIOMA. The user is advised to use the control tissues available from BioGenex for your Quality Control purpose. Refer to the appropriate detection system package inserts for guidance on general quality control procedures.

Troubleshooting

Refer to the troubleshooting section in the package inserts of BioGenex Detection Systems (or other equivalent detection systems) for remedial actions on detection system related issues, or contact BioGenex Technical Support Department at 1-800-421-4149 or your local distributor to report unusual staining.

Expected Results

This antibody stains nuclear in positive cells in formalin-fixed, paraffinembedded tissue sections. Interpretation of the staining result is solely the



responsibility of the user. Any experimental result should be confirmed by a medically established diagnostic product or procedure.

Limitations of the Procedure

Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can also cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue may cause variations in results (Nadji and Morales, 1983). Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive with horseradish peroxidase systems (Omata et al, 1980). Improper counterstaining and mounting may compromise the interpretation of results.

Performance Characteristics

BioGenex has conducted studies to evaluate the performance of the antibody with BioGenex detection systems and accessories. The antibodies have been found to be sensitive and show specific binding to the antigen of interest with minimal to no binding to non-specific tissues or cells. BioGenex antibodies have shown reproducible and consistent results when used within a single run, between runs, between lots and wherever applicable between manual and automated runs. The products have been determined to be stable for the periods specified on the labels either by standard real time or accelerated methods. BioGenex ensures product quality through 100% quality control for all products released and through surveillance programs.

Bibliography

- 1. Barberis MCP, et al: "Calretinin. A selective marker of normal and neoplastic mesothelial cells in serous effusions". ActaCytol 1997;
- Versnel MA, et al: "Establishment of human malignant mesothelioma cell lines". Int J Caner 1989;44:256-60.
- Versnel MA, et al." Characterization of three human malignant mesothelioma cell lines". Cancer Genet Cytogenet1989;42:115-28
- Tong Ma, et al: "A subpopulation of dorsal lateral/caudal ganglionic eminence-derived neocortical interneurons expresses the transcription factor Sp8". Cerebral cortex (New York, N.Y.: 1991).22
- Center for Disease Control. Decontamination of Laboratory Sink Drains to Remove Azide Salts. Center for Disease Control Manual Guide--Safety Management, No. CDC-22, Atlanta, Georgia. April 30,
- Kiernan JA. Histological and Histochemical Methods: Theory and Practice. New York: Pergamon Press 1981.
- Nadji M, Morales AR. Immunoperoxidase, part 1: the techniques and its pitfall. Lab Med 1983; 14:767-770.
- Omata M, Liew CT, Ashcavai M, Peters Rl. Nonimmunologic binding of horseradish peroxidase to hepatitis B surface antigen. A possible source of error in immunohistochemistry. Am J ClinPathol. May, 1980;73(5):626-632.
- 9. U.S. Congress. Clinical Laboratory Improvement Amendments of 1988: Final Rule, 57 FR 7163, February 28, 1992.
- 10. National Institute for Occupational Safety and Health, (NIOSH), Rockville, MD. Explosive azide hazard, Publication No. 78-127, 1976.

Doc. No. 932-747N-EN Rev. D Release Date: 18-Jul-2016

^{**} Pretreatment times will vary based on individual microwave power

^{***}For Closed system automation (Xmatrx-Elite, Xmatrx-Ultra & i6000 Elite Dx)

⁻ Please refer to the factory protocols provided with the instrument.